
Rules of
Department of Natural Resources
Division 60—Public Drinking Water Program
Chapter 5—Laboratory and Analytical
Requirements

Title	Page
10 CSR 60-5.010 Accepted and Alternate Procedures for Analyses	3
10 CSR 60-5.020 Laboratory Certification	15

**Title 10—DEPARTMENT OF
NATURAL RESOURCES
Division 60—Public Drinking Water
Program
Chapter 5—Laboratory and Analytical
Requirements**

**10 CSR 60-5.010 Accepted and Alternate
Procedures for Analyses**

PURPOSE: This rule lists manuals containing acceptable analysis procedures for determination of contaminant levels.

PUBLISHER'S NOTE: The publication of the full text of the material that the adopting agency has incorporated by reference in this rule would be unduly cumbersome or expensive. Therefore, the full text of that material will be made available to any interested person at both the Office of the Secretary of State and the office of the adopting agency, pursuant to section 536.031.4, RSMo. Such material will be provided at the cost established by state law.

(1) Inorganic and Radiological Contaminants. Unless substitute methods are approved by the department, analysis shall be conducted in accordance with the methods listed in the following table. All analytical methods listed in this rule are incorporated by reference. Criteria for analyzing arsenic, barium, beryllium, cadmium, calcium, chromium, copper, lead, nickel, selenium, sodium, and thallium with digestion or directly without digestion, and other analytical test procedures are contained in "Technical Notes on Drinking Water Methods," EPA-600/R-94-173, October 1994. This document also contains approved analytical test methods which remain available for compliance monitoring until July 1, 1996. These methods will not be available for use after July 1, 1996. This document is available from the National Technical Information Service, NTIS PB95-104766, U.S. Department of Commerce, 5285 Port Royal Road, Springfield, VA 22161. The toll-free number is 800-553-6847.



Contaminant	Methodology (if appropriate)	Method (if appropriate)	Reference
Aluminum	Inductively Coupled Plasma—Emission Spectroscopy	200.7	22
		3120 B	17
	Inductively Coupled Plasma—Mass Spectrometry Atomic Absorption—Platform Technique	200.8	22
		200.9	22
		3113 B 3111 D	17 17
Antimony	Atomic Absorption—Gaseous Hydride	D3697-92	19
	Atomic Absorption—Graphite Furnace	3113 B	17
	Inductively Coupled Plasma—Mass Spectrometry	200.8	22
	Atomic Absorption—Platform Technique	200.9	22
Arsenic	Atomic Absorption—Graphite Furnace	D2972-93C	19
		3113 B	17
	Atomic Absorption—Gaseous Hydride	D2972-93B	19
		3114 B	17
	Atomic Absorption—Platform	200.9	22
	Inductively Coupled Plasma—Emission Spectroscopy	200.7	22
		3120 B	17
200.8		22	
Asbestos	Transmission Electron Microscopy	100.1	7
		100.2	27
Barium	Inductively Coupled Plasma—Mass Spectrometry	200.8	22
		3113 B	17
	Atomic Absorption—Graphite Furnace	3111 D	17
	Atomic Absorption—Direct Aspiration	200.7	22
	Inductively Coupled Plasma—Emission Spectroscopy	3120 B	17
Beryllium	Atomic Absorption—Graphite Furnace	D3645-93B	19
		3113 B	17
	Atomic Absorption—Platform	200.9	22
	Inductively Coupled Plasma—Emission Spectroscopy	3120 B	17
		200.7	22
		200.8	22
Cadmium	Atomic Absorption—Graphite Furnace	3113 B	17
	Atomic Absorption—Platform	200.9	22
	Inductively Coupled Plasma—Emission Spectroscopy	200.7	22
	Inductively Coupled Plasma—Mass Spectrometry	200.8	22
Chloride		300.0	25
		D4327-91	19
		4500-Cl ⁻ D	17
Chromium	Atomic Absorption—Furnace Technique	3113 B	17
		200.9	22
	Inductively Coupled Plasma—Emission Spectroscopy	200.7	22
		3120 B	17
		200.8	22
Inductively Coupled Plasma—Mass Spectrometry			
Color		2120 B	17
Copper	Atomic Absorption—Furnace Technique	D1688-90C	19
		3113 B	17
		200.9	22
Atomic Absorption—Platform			

Contaminant	Methodology (if appropriate)	Method (if appropriate)	Reference
Copper (cont.)	Atomic Absorption—Direct Aspiration	D1688-90A	19
	Inductively Coupled Plasma	3111 B	17
		200.7	22
	Inductively Coupled Plasma—Mass Spectrometry	3120 B	17
		200.8	22
Cyanide	Manual distillation followed by—	4500-CN ⁻ C	17
	1. Amenable Spectrophotometric	D2036-91B	19
		4500-CN ⁻ G	17
	2. Spectrophotometric—Manual	D2036-91A	19
		4500-CN ⁻ E	17
		1-3300-85	24
	3. Spectrophotometric— Manual—Semiautomated	335.4	25
	4. Selective Electrode	4500-CN ⁻ F	17
Fluoride	Colorimetric SPADNS, with distillation	4500-F B&D	17
	Ion Chromatography	300.0	25
		D4327-91	19
		4110B	17
	Manual Electrode	D1179-93B	19
		4500-F ⁻ C	17
	Automated Alizarin Fluoride Blue, with distillation (complexone)	4500-F E	17
		129-71W	9
Automated Ion Selective Electrode	380-75WE	10	
Foaming Agents		5540 C	17
Iron		200.7	22
		200.9	22
		3120 B	17
		3111 B	17
		3113 B	17
Lead	Atomic Absorption—Furnace Technique	D3559-90D	19
		3113 B	17
	Inductively Coupled Plasma—Mass Spectrometry	200.8	22
	Atomic Absorption—Platform Furnace	200.9	22
Manganese		200.7	22
		200.8	22
		200.9	22
		3120 B	17
		3111 B	17
		3113 B	17
Mercury	Manual cold vapor technique	245.1	22
		D3223-91	19
		3112 B	17
	Automated cold vapor technique	245.2	1
	Inductively Coupled Plasma—Mass Spectrometry	200.8	22
Nickel	Atomic Absorption—Direct Aspiration	3111 B	17
	Atomic Absorption—Platform Technique	200.9	22



Contaminant	Methodology (if appropriate)	Method (if appropriate)	Reference
Nickel (cont.)	Inductively Coupled Plasma—Emission Spectroscopy	200.7	22
		3120 B	17
	Inductively Coupled Plasma—Mass Spectrometry Atomic Absorption—Graphite Furnace	200.8 3113 B	22 17
Nitrate	Manual Cadmium Reduction	D3867-90B	19
		4500-NO ₃ ⁻ E	17
	Automated Cadmium Reduction	353.2	25
		D3867-90A	19
		4500-NO ₃ ⁻ F	17
	Ion Selective Electrode	4500-NO ₃ ⁻ D	17
		601	26
	Ion Chromatography	300.0	25
		B1011	8
4110 B D4327-91		17 19	
Nitrite	Spectrophotometric Automated Cadmium Reduction	4500-NO ₃ ⁻ B	17
		353.2	25
	Manual Cadmium Reduction	D3867-90A	19
		4500-NO ₃ ⁻ F	17
		D3867-90B	19
	Ion Chromatography	4500-NO ₃ ⁻ E	17
		300.0	25
		B1011	8
		D4327-91 4110 B	19 17
Odor		2150 B	17
Operational Monitoring— General			17
Operational Monitoring— pH	pH Value	150.1	1
		150.2	1
		D1293-84	19
		4500-H ⁺ B	17
Operational Monitoring— Residual Disinfectant Monitoring— Chlorine Dioxide	Amperometric Titration	4500-ClO ₂ C	17
		4500-ClO ₂ E	17
		DPD Method	4500-ClO ₂ D
Operational Monitoring— Residual Disinfectant Monitoring— Free Chlorine	Amperometric Titration DPD Ferrous Titrimetric DPD Colorimetric Syringaldazine (FACTS)	4500-Cl D	17
		4500-Cl F	17
		4500-Cl G	17
		4500-Cl H	17
		4500-Cl I	17
Combined Chlorine	Amperometric Titration Amperometric Titration—low level measurement DPD Ferrous Titrimetric DPD Colorimetric Iodometric Electrode	4500-Cl D	17
		4500-Cl E	17
		4500-Cl F	17
		4500-Cl G	17
		4500-Cl I	17

Contaminant	Methodology (if appropriate)	Method (if appropriate)	Reference	
Operational Monitoring— Lead and Copper Regulation	pH Electrometric	150.1	1	
		150.2	1	
		D1293-84	19	
	Conductivity	4500-H ⁺ B	17	
		D1125-91A	19	
	Calcium (EDTA Titrimetric)	2510 B	17	
		D511-93A	19	
		3500-Ca D	17	
	(Inductively Coupled Plasma)	200.7	22	
	(Atomic Absorption—Direct Aspiration)	3120 B	17	
		D511-93B	19	
		3111 B	17	
	Alkalinity (Titrimetric)	D1067-92B	19	
		2320 B	17	
	(Electrometric Titration)	I-1030-85	24	
	Orthophosphate (unfiltered, no digestion or hydrolysis) (Colorimetric, automated ascorbic acid)	365.1	25	
			4500-P F	17
		(Colorimetric, ascorbic acid single reagent)	D515-88A	19
			4500-P E	17
		(Colorimetric, phosphomolybdate, automated—segmented flow, automated discrete)	I-1601-85	24
			I-2601-90	24
			I-2598-85	24
		(Ion Chromatography)	300.0	25
		Silica (Colorimetric, molybdate blue, automated—segmented flow)	D4327-91	19
			4110	17
	I-1700-85		24	
	I-2700-85		24	
D859-88	19			
4500-Si D	17			
4500-Si E	17			
(Automated method for molybdate—reatine silica)	4500-Si F	17		
(Inductively Coupled Plasma)	200.7	22		
Operational Monitoring— Residual Disinfectant Monitoring— Ozone	Indigo	3120 B	17	
		4500-O B	17	
Operational Monitoring— Temperature	Thermometric	2550 B	17	
Radionuclides	Examination of Water & Wastewater for Radioactivity		11	
			12	
			13	
		700	3	



Contaminant	Methodology (if appropriate)	Method (if appropriate)	Reference	
Radionuclides (cont.)	Gamma Spectrometry in Water	D2459	2	
	Microquantities of Uranium in Water by Fluorometry	D2907	2	
Secondary Contaminants			3	
Selenium	Atomic Absorption—Hydride Generation	D3859-93A	19	
		3114 B	17	
	Atomic Absorption—Graphite Furnace	D3859-93	19	
		3113 B	17	
	Atomic Absorption—Platform	200.9	22	
	Inductively Coupled Plasma—Mass Spectrometry	200.8	22	
	Silver	Atomic Absorption—Graphite Furnace	I-3720-85	24
		Inductively Coupled Plasma—Emission Spectroscopy	200.7	22
Inductively Coupled Plasma—Mass Spectrometry		200.8	22	
Atomic Absorption—Platform Technique		200.9	22	
Sulfate		3120 B	17	
		3111 B	17	
		3113 B	17	
	Colorimetric—Methylthymol blue	375.2	25	
	Gravimetric	4500-SO ₄ ²⁻ C, D	17	
	Turbidimetric	4500-SO ₄ ²⁻ F	17	
Ion Chromatography		300.0	25	
		4110	17	
		D4327-91	19	
Thallium	Atomic Absorption—Platform Technique	200.9	22	
	Inductively Coupled Plasma—Mass Spectrometry	200.8	22	
Total Dissolved Solids		2540 C	17	
Turbidity	Nephelometric	2130 B	17	
		180.1	25	
	Great Lakes Instruments	Method 2	18	
Sodium	Inductively Coupled Plasma	200.7	22	
	Atomic Absorption—Direct Aspiration	3111 B	17	
Zinc		200.8	22	
		3111 B	17	

(A) References for analytical methods in 10 CSR 60-5.010(1):

1. "Methods of Chemical Analysis of Water and Wastes," EPA Environmental Monitoring and Support Laboratory, Cincinnati, OH 45268 (EPA-600/4-79-020), March 1983. Available from National Technical Information Service, PB84-128677. Methods 150.1, 150.2 and 246.2 are also available from U.S. EPA, EMSL, Cincinnati, OH 45268.

2. *Annual Book of ASTM Standards*, Vols. 11.01 and 11.02, 1991, American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.

3. *Standard Methods for the Examination of Water and Wastewater*, 16th edition, American Public Health Association, American Water Works Association, Water Pollution Control Federation, 1985.

4. "Methods for Determination of Inorganic Substances in Water and Fluvial Sediments," *Techniques of Water-Resources Investigations of the U.S. Geological Survey Books*, Book 5, Chapter A1, Third Edition, 1989. Available at Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

5. "Orion Guide to Water and Wastewater Analysis." Form WeWWG/5880, p. 5, 1985. Orion Research, Inc., Cambridge, MA 02139.

6. 200.7A "Inductively Coupled Plasma Atomic Emission Analysis of Drinking Water," Appendix to Method 200.7, March 1987, U.S. EPA, Environmental Monitoring and Support Laboratory, Cincinnati, OH 45268.

7. "Analytical Method for Determination of Asbestos Fibers in Water," Method 100.1, EPA-600/4-83-043, September 1983. Available from National Technical Information Service, PB83-260471.
8. "Waters Test Method for the Determination of Nitrite/Nitrate in Water Using Single Column Ion Chromatography," Method B-1011, Millipore Corporation, Waters Chromatography Division, 34 Maple Street, Milford, MA 01757.
9. "Fluoride in Water and Wastewater," Industrial Method 129-71W, Technicon Industrial Systems, Tarrytown, NY 10591, December 1972.
10. "Fluoride in Water and Wastewater," Industrial Method No. 380-75WE, Technicon Industrial Systems, Tarrytown, NY 10591, February 1976.
11. "Radiochemical Methodology for Drinking Water," Environmental Monitoring Support Laboratory, EPA-600/4-75-008, U.S. EPA, Cincinnati, OH 45268.
12. "Procedures for Radiochemical Analysis of Nuclear Reactor Aqueous Solutions," H.L. Krieger and S. Gold, EPA-R4-730014, U.S. EPA, Cincinnati, OH, May 1973.
13. *HASL Procedure Manual*, edited by John H. Harley, HASL 300, ERDA Health and Safety Laboratory, New York, NY, 1973.
14. "Determination of and Trace Elements in Water and Wastes by Inductively Coupled Plasma-Mass Spectrometry," Method 200.8, version 4.3, August 1990, EPA, Environmental Monitoring and Systems Laboratory, Cincinnati, OH 45268. Available from ORD Publication, CERL, EPA, Cincinnati, OH 45268.
15. "Determination of Metals and Trace Elements by Stabilized Temperature Graphite Furnace Atomic Absorption Spectrometry," Method 200.9, version 1.1, August 1990, EPA, Environmental Monitoring and Systems Laboratory, Cincinnati, OH 45268.
16. "Determination of Ozone in Water by the Indigo Method; A Submitted Standard Method," *Ozone Science and Engineering*, Volume 4, pages 169-176, Pergamon Press Ltd., 1982.
17. *Standard Methods for the Examination of Water and Wastewater*, 18th edition, American Public Health Association, American Water Works Association, Water Pollution Control Federation, 1992.
18. "Turbidity," GLI Method 2, November 2, 1992, Great Lakes Instruments, Inc., 8855 North 55 Street, Milwaukee, WI 53223.
19. *Annual Book of ASTM Standards*, Vols. 11.01 and 11.02, 1994, American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.
20. "Determination of Metals and Trace Elements in Water and Wastes by Inductively Coupled Plasma-Mass Spectrometry" Method 200.8, August 1990, Revision 3.2 EPA EMSL. Available from U.S. EPA, EMSL Cincinnati, OH 45268.
21. "Determination of Inorganic Ions in Water by Ion Chromatography" Method 300.8, December 1989, U.S. EPA EMSL. Available from U.S. EPA, EMSL, Cincinnati, OH 45268.
22. "Methods for the Determination of Metals in Environmental Samples—Supplement I, EPA-600/R-94-111, May 1994." Available from National Technical Information Service (NTIS) NTIS PB 94-184942, U.S. Department of Commerce, 5285 Port Royal Road, Springfield, VA 22161. The toll free number is (800) 553-6847.
23. "Method 300. Determination of Inorganic Anions in Water by Ion Chromatography." Inorganic Chemistry Branch, Environmental Monitoring Systems Laboratory, August 1991.
24. Available from the Books and Open-File Reports Section, United States Geological Survey, Federal Center, Box 25425, Denver, CO 80225-0425.
25. "Methods for the Determination of Inorganic Substances in Environmental Samples," EPA-600/R-93-100, August 1993. Available from National Technical Information Service, PB94-121811.
26. The procedure shall be done in accordance with the Technical Bulletin 601, "Standard Method of Test for Nitrate in Drinking Water," July 1994, PN 221890-001, Analytical Technology, Inc. Available from ATI, Orion, 529 Main Street, Boston, MA 02129.
27. "Determination of Asbestos Structure over 10- μ m in Length in Drinking Water," Method 100.2, EPA-600/R-94-134, June 1994. Available from NTIS, PB94-201902.

(2) Organic Contaminants. Unless substitute methods are approved, the following table includes acceptable analysis procedures for organic contaminants:

<u>Contaminant</u>	<u>Method</u>	<u>Contaminant</u>	<u>Method</u>
		Dinoseb	515.1 515.2 555
		Diquat	549.1
		Endothall	548.1
		Endrin	505 508 525.2 508.1
2,3,7,8-TCDD (Dioxin)	1613	Ethylene dibromide (EDB)	504.1 551
2,4-D	515.1 515.2 555	Glyphosate	547 6651
2,4,5-TP (Silvex)	555	Heptachlor	505 508 525.2 508.1
515.1515.2,	531.1	Heptachlor epoxide	505 508 525.2
3-Hydrodroxycarbofuran	6610	Hexachlorobenzene	505 508 525.2 508.1
Alachlor	505 507 525.2 508.1	Hexachlorocyclopentadiene	505 525.2 508 508.1
Aldicarb	531.1 6610	Lindane	505 508 525.2 508.1
Aldicarb sulfoxide	531.1 6610	Methomyl	531.1 6610
Aldicarb sulfone	531.1 6610	Methoxychlor	505 525.2 508.1
Aldrin	505 508 525.2 508.1	Metolachlor	507 508.1 525.2
Atrazine	505 507 525.2 508.1	Metribuzin	507 508.1 525.2
Benzo(a)pyrene	525.2 550 550.1	Oxamyl (vydate)	531.1 6610
Butachlor	507 525.2	Pentachlorophenol	515.1 515.2 525.2 555
Carbaryl	531.1 6610	Picloram	515.1 515.2 555
Carbofuran	531.1 6610	Polychlorinated biphenyls	505 (as Aroclors) 508 (as Aroclors) 508A (as deca-chlorobiphenyl)
Chlordane	505 508 525.2 508.1		
Dalapon	515.1 552.1		
Di(2-ethylhexyl)adipate	506 525.2		
Di(2-ethylhexyl)phthalate	506 525.2		
Dibromochloropropane (DBCP)	504.1 551		
Dicamba	515.1 515.2 555		
Dieldrin	505 508 508.1 525.2		
		Propachlor	508 508.1 525.2
		Simazine	505 507 508.1 525.2
		Toxaphene	505 508 525.2
		Total Trihalomethanes	502.2 524.2 551
		Volatile Organic Chemicals (regulated and unregulated)	502.2 524.2
		Footnotes	
		1) A nitrogen-phosphorous detector should be substituted for the electron capture detector in Method 505 (or another approved method should be used) to determine alachlor, atrazine and simazine, if lower detection limits are required.	
		2) PCBs are qualitatively identified as Aroclors and measured for compliance purposes as decachlorobiphenyl. Each system which monitors for PCBs shall analyze each sample using either Method 505 or Method 508.	
		3) Analyses of total trihalomethanes shall be conducted in accordance with these methods and "Technical Notes on Drinking Water Methods," EPA-600/R94-173, October 1994, which is available at NTIS, PB95-104766.	
		4) In addition to Methods 502.2 and 524.2, analysis for bromodichloromethane, bromoform,* chlorodibromomethane, chloroform, carbon tetrachloride, tetrachlorethylene, 1,1,1-trichloroethane, and trichloroethylene may also be conducted by EPA Method 551. Analysis for 1,2,3-trichloropropane may be conducted by Methods 502.1, 524.2 and 504.1.	
		References for analytical methods in 10 CSR 60-5.010(2): Methods 502.2, 505, 507, 508, 508A, 515.1 and 531.1 are in "Methods for the Determination of Organic Compounds in Drinking Water," EPA-600/4-88-039, December 1988, revised July 1991. Methods 506, 547, 550, 550.1 and 551 are in "Methods for the Determination of Organic Compounds in Drinking Water—Supplement I," EPA-600-4-90-020, July 1990. Methods 515.2, 524.2, 548.1, 549.1, 552.1 and 555 are in "Methods for the Determination of Organic Compounds in Drinking Water—Supplement II," EPA-600/R-92-129, August 1992. Method 1613 is titled "Tetra-through Octa-Chlorinated Dioxins and Furans by Isotope-Dilution HRGC/HRMS," EPA-821-	

- B-94-005, October 1994. These documents are available from National Technical Information Service (NTIS) NTIS PB91-231480, PB91-146027, PB92-207703, and PB95-104774, U.S. Department of Commerce, 5285 Port Royal Road, Springfield, VA 22161. The toll free number is (800) 553-6847. Method 6651 shall be followed in accordance with the 18th edition of "Standard Methods for the Examination of Water and Wastewater," 1992. Available from the American Public Health Association, 1015 Fifteenth Street NW, Washington, DC 20005. Method 6610 shall be followed in accordance with the "Supplement to the 18th Edition of Standard Methods for the Examination of Water and Wastewater," 1994. Available from the American Public Health Association, 1015 Fifteenth Street NW, Washington, DC 20005. EPA Methods 504.1, 508.1 and 525.2 are available from U.S. EPA EMSL, Cincinnati, OH 45268. The phone number is (513) 569-7586. Other analytical test procedures are contained in Technical Notes on Drinking Water Methods, EPA-600/R-94-173, October 1994, NTIS PB95-104766.
- 502.1 Volatile halogenated organic chemicals in water by purge and trap gas chromatography
- 502.2 Volatile organic compounds in water by purge and trap capillary column gas chromatography with photoionization and electrolytic conductivity detectors in series
- 503.1 Volatile aromatic and unsaturated organic compounds in water by purge and trap gas chromatography
- 504.1 1,2-Dibromoethane (EDB), 1,2-Dibromo-3-chloropropane (DBCP), and 1,2,3-Trichloropropane (123TCP) in Water by Microextraction and Gas Chromatography
- 505 Analysis of organohalide pesticides and commercial polychlorinated biphenyl products (Aroclors) in water by microextraction and gas chromatography
- 506 Determination of phthalate and adipate esters in drinking water by liquid-liquid extraction or liquid-solid extraction and gas chromatography with photoionization detection
- 507 Determination of nitrogen- and phosphorus-containing pesticides in groundwater by gas chromatography with a nitrogen-phosphorus detector
- 508 Determination of chlorinated pesticides in water by gas chromatography with an electron capture detector
- 508.1 Determination of chlorinated pesticides, herbicides, and organohalides by liquid-solid extraction and electron capture gas chromatography
- 508A Screening for polychlorinated biphenyls by perchlorination and gas chromatography (for quantification if detected with Method 505 or 508)
- 515.1 Determination of chlorinated acids in water by gas chromatography with an electron capture detector, revision 5.0, May 1991
- 524.1 Measurement of purgeable organic compounds in water by purged column gas chromatography/mass spectrophotometry
- 524.2 Volatile organic chemicals in water by purge and trap capillary column gas chromatography/mass spectrophotometry
- 525.2 Determination of organic compounds in drinking water by liquid-solid extraction in capillary column gas chromatography/mass spectrometry
- 531.1 Measurement of N-methyl carbamoyloximes and N-methyl carbamates in water by direct aqueous injection HPLC with post-column derivatization
- 547 Analysis of glyphosate in drinking water by direct-aqueous-injection HPLC, with post-column derivatization
- 548 Determination of endothall in aqueous samples
- 549.1 Determination of diquat and paraquat in drinking water by liquid-solid extraction and high performance liquid chromatography with ultraviolet detection
- 550 Determination of polycyclic aromatic hydrocarbons in drinking water by liquid-liquid extraction and HPLC with coupled ultraviolet and fluorescence detection
- 550.1 Determination of polycyclic aromatic hydrocarbons in drinking water by liquid-solid extraction and HPLC with coupled ultraviolet and fluorescence detection
- 552 Determination of polycyclic aromatic hydrocarbons in drinking water by liquid-solid extraction and HPLC with coupled ultraviolet and fluorescence detection
- 551 Determination of chlorination disinfection byproducts and chlorinated solvents in drinking water by liquid-liquid extraction and gas chromatography with electron-capture detection
- 555 Determination of chlorinated acids in water by high performance liquid chromatography with a photodiode array ultraviolet detector
- 1613 "Tetra-through Octa-Chlorinated Dioxins and Furans by Isotope Dilution." This method is available from U.S. EPA-OST, Sample Control Center, P.O. Box 1407, Alexandria, VA 22313.
- 6610 Carbamate pesticides
- 6651 Glyphosate herbicide
- (3) Microbiological Contaminants. Unless substitute methods are approved, this section (3) lists acceptable analysis procedures for microbiological contaminants.

Contaminant

Approved Manual or Procedure

(A) *Escherichia coli*.
the (*E. coli*)

Public water systems must conduct analysis of *Escherichia coli* in accordance with one (1) of following analytical methods:

1. EC medium supplemented with fifty (50) μ /ml of 4-methylumbelliferyl-beta-D-glucuronide (MUG) (final concentration). EC medium is described in *Standard Methods for the Examination of Water and Wastewater*, 1992, American Public Health Association, 18th edition, Method 9221E, p. 9-52, paragraph 1a. MUG may be added to EC medium before autoclaving. EC medium supplemented with fifty (50) μ /ml of MUG is commercially available. At least ten (10) ml of EC medium supplemented with MUG must be used. The inner inverted fermentation tube may be omitted. The procedure for transferring a total coliform-positive culture to EC medium supplemented with MUG shall be as specified in 10 CSR 60-5.010(1)(B)2. for transferring a total coliform-positive culture to EC medium. Observe fluorescence with an ultraviolet light (366 nm) in the dark after incubating tube at 44.5 ± 0.2 degrees Celsius for 24 ± 2 hours;

2. Nutrient agar supplemented with 100 μ /ml MUG (final concentration). Nutrient agar is described in *Standard Methods for the Examination of Water and Wastewater*, 1992, American Public Health Association, 18th edition, p. 9-47 to 9-48. This test is used to determine if a total coliform-positive sample, as determined by the membrane filter technique (MFT) or any other method in which a membrane filter is used, contains *E. coli*. Transfer the membrane filter containing a total coliform colony(ies) to nutrient agar supplemented with 100 μ /ml (final concentration) MUG. After incubating the agar plate at thirty-five degrees Celsius (35°C) for four (4) hours, observe the colony(ies) under ultraviolet light (366 nm) in the dark for fluorescence. If fluorescence is visible, *E. coli* are present;

3. Minimal Medium ONPG-MUG (MMO-MUG) Test, as set forth in the article “National Field Evaluation of a Defined Substrate Method for the Simultaneous Detection of Total Coliforms and *Escherichia coli* from Drinking Water: Comparison with Presence-Absence Techniques” (Edberg et al.), *Applied and Environmental Microbiology*, Volume 55, pp. 1003-1008, April 1989. (Note: The Autoanalysis Colilert System is an MMO-MUG test.) If the MMO-MUG test is total coliform-positive after a 24-hour incubation, test the medium for fluorescence with a 366-nm ultraviolet light (preferably with a 6-watt lamp) in the dark. If fluorescence is observed, the sample is *E. coli*-positive. If fluorescence is questionable (cannot be definitively read) after 24 hours incubation, incubate the culture for an additional four (4) hours (but not to exceed 28 hours total). And again test the medium for fluorescence. The MMO-MUG Test with hepes buffer in lieu of phosphate buffer is the only approved formulation for the detection of *E. coli*;

4. As an option to paragraph (3)(A)3. of this rule, a system with a total coliform-positive, MUG-negative, MMO-MUG test may further analyze the culture for the presence of *E. coli* by transferring a 0.1 ml 28-hour MMO-MUG culture to EC Medium + MUG. Observation of the results are described in 40 CFR 141.21(f)(6)(i); or

5. The Colisure test. A description of the Colisure test may be obtained from the Millipore Corporation, Technical Services Department, 80 Ashby Road, Bedford, MA 01730.

(B) Fecal Coliform.

1. Public water systems must conduct fecal coliform analysis in accordance with the following procedure: When the multiple-tube fermentation (MTF) technique or presence-absence (P-A) coliform test is used to test for total coliform, shake the lactose-positive presumptive tube or P-A vigorously and transfer the growth with a sterile three millimeter (3 mm) loop or sterile applicator stick into brilliant green lactose bile broth and EC medium to determine the presence of total and fecal coliforms, respectively. For EPA-approved analytical methods which use a membrane filter, transfer the total coliform-positive culture by one (1) of the following methods: Remove the membrane containing the total coliform colonies from the substrate with a sterile forceps and carefully curl and insert the membrane into a tube of EC medium (the laboratory may first remove a small portion of selected colonies for verification), swab the entire membrane filter surface with a sterile cotton swab and transfer the inoculum to EC medium (do not leave the cotton swab in the EC medium), or inoculate individual total coliform-positive colonies into EC medium. Gently shake the inoculated tubes of EC medium to insure adequate mixing and incubate in a waterbath at 44.5 ± 0.2 degrees Celsius for 24 ± 2 hours. Gas production of any amount in the inner fermentation tube of the EC medium indicates a positive fecal coliform test. The preparation of EC medium is described in *Standard Methods for the Examination of Water and Wastewater*, 1992, American Public Health Association, 18th edition, Method 9221E, p. 9-2, paragraph 1a. Public water systems need only determine the presence or absence of fecal coliforms; a determination of fecal coliform density is not required. Only this method for fecal coliform is allowed for compliance with 10 CSR 60-4.020(5); or

2. *Standard Methods for the Examination of Water and Wastewater*, 1992, American Public Health Association, 18th edition.

A. Fecal coliform most probable number (MPN) procedures, method 9221E, pages 9-52 to 9-53 (Note: A-1 Broth may be held up to three (3) months in a tightly closed screwcap tube at four degrees Celsius (4°C);

B. Fecal coliform membrane filter procedure, method 9222D, pages 9-60 to 9-61.

Contaminant	Approved Manual or Procedure
(C) Heterotrophic Bacteria.	<i>Standard Methods for the Examination of Water and Wastewater</i> , 1992, American Public Health Association, 18th edition, pour plate method, method 9215B. The time from sample collection to initiation of analysis may not exceed twenty-four (24) hours.
(D) Total Coliform.	<p>1. The standard sample volume required for total coliform analysis, regardless of analytical method used, is one hundred milliliters (100 ml). The time from sample collection to initiation of analysis may not exceed forty-eight (48) hours. If the laboratory analyzes samples after thirty (30) hours and up to forty-eight (48) hours from sample collection, the laboratory shall indicate on the report of the analysis results that the data may be invalid because of excessive delay before sample processing.</p> <p>2. <i>Standard Methods for the Examination of Water and Wastewater</i>, 1992, American Public Health Association, 18th edition—</p> <p>A. Fermentation technique, method 9221A, B.</p> <p>(I) Lactose broth, as commercially available, may be used in lieu of lauryl tryptose broth, if the system conducts at least twenty-five (25) parallel tests between this medium and lauryl tryptose broth using the water normally tested, and this comparison demonstrates that the false-positive rate for total coliforms, using lactose broth, is less than ten percent (10%).</p> <p>(II) If inverted tubes are used to detect gas production, the media should cover these tubes at least one-half (1/2) to two-thirds (2/3) after the sample is added.</p> <p>(III) No requirement exists to run the completed phase on ten percent (10%) of all total coliform-positive confirmed tubes.</p> <p>B. Membrane filter (MF) technique, method 9222A, B, C.</p> <p>C. Presence-absence (P-A) coliform test, method 9221D.</p> <p>(I) Six-times formulation strength may be used if the medium is filter-sterilized rather than autoclaved.</p> <p>(II) No requirement exists to run the completed phase on ten percent (10%) of all total coliform-positive confirmed tubes.</p> <p>D. ONPG-MUG Test (also known as the Autoanalysis Colilert System, method 9223).</p> <p>E. Colisure test. The Colisure test must be incubated for twenty-eight (28) hours before examining the results. If an examination of the results at twenty-eight (28) hours is not convenient, then results may be examined at any time between twenty-eight (28) and forty-eight (48) hours. A description of the Colisure test may be obtained from the Millipore Corporation, Technical Service Department, 80 Ashby Road, Bedford, MA 01730.</p>

(4) Sample collection for the contaminants listed in this rule must be conducted using the sample preservation, container and maximum holding time procedures specified in the following table. All other samples for contaminants in 10 CSR 60-5.010 shall be collected in accordance with procedures contained in the appropriate analytical method.

Contaminant	Preservative	Container	Holding Time
Antimony	Concentrated HNO ₃ to pH < 2	P or G	6 months
Asbestos	Cool to 4°C	P or G	
Barium	Concentrated HNO ₃ to pH < 2	P or G	6 months
Beryllium	Concentrated HNO ₃ to pH < 2	P or G	6 months
Cadmium	Concentrated HNO ₃ to pH < 2	P or G	6 months
Chromium	Concentrated HNO ₃ to pH < 2	P or G	6 months
Copper			
Preserved	Concentrated HNO ₃ to pH < 2	P or G	6 months
Unpreserved	NONE	P or G	14 days
Cyanide	Cool to 4°C, NaOH to pH > 12	P or G	14 days
Fluoride	NONE	P or G	1 month
Lead			
Preserved	Concentrated HNO ₃ to pH < 2	P or G	6 months
Unpreserved	NONE	P or G	14 days
Mercury	Concentrated HNO ₃ to pH < 2	P or G	28 days

Contaminant	Preservative	Container	Holding Time
Nickel	Concentrated HNO ₃ to pH < 2	P or G	6 months
Nitrate			
Chlorinated	Cool to 4°C	P or G	28 days
Nonchlorinated	Concentrated H ₂ SO ₄ to pH < 2	P or G	14 days
Nitrite	Cool to 4°C	P or G	48 hours
Selenium	Concentrated HNO ₃ to pH < 2	P or G	6 months
Thallium	Concentrated HNO ₃ to pH < 2	P or G	6 months

(A) If HNO₃ cannot be used because of shipping restrictions, sample may be initially preserved by icing and immediately shipping to the laboratory. Upon receipt in the laboratory, the sample must be acidified with concentrated HNO₃ to pH < 2 and held for sixteen (16) hours before analysis. At time of analysis, sample container should be thoroughly rinsed with 1:1 HNO₃; washings should be added to the sample.

(B) P = Plastic, hard or soft.

(C) G = Glass, hard or soft.

(D) In all cases samples should be analyzed as soon after collection as possible.

(E) For cyanide see method(s) for the information for preservation.

(F) The system shall take each sample at the same sampling point unless conditions make another sampling point more representative of each source or treatment plant.

(G) If a system draws water from more than one (1) source and the sources are combined before distribution, the system must sample at an entry point to the distribution system during periods of normal operating conditions (that is, when water is representative of all sources being used).

(5) The department may reduce the total number of samples a system must analyze by allowing the use of compositing. Composite samples from a maximum of five (5) sampling points are allowed provided that the detection limit of the method used for analysis is less than one-fifth (1/5) of the MCL. For a lower number of composited samples the allowable relationship between the detection limit and MCL will change proportionally. Compositing of samples must be done in the laboratory and the composite sample must be analyzed within fourteen (14) days of collection. If the population served by the system is greater than thirty-three hundred (>3,300) persons, then compositing is permitted only at sampling points within a single system. In systems serving less than or equal to thirty-three hundred (≤3,300) persons, the department may permit compositing among different systems provided the five (5)-sample limit is maintained.

(A) Compositing of samples may be allowed for inorganic chemicals (IOCs) listed

in 10 CSR 60-4.030(1) and synthetic organic chemicals (SOCs) listed in 10 CSR 60-4.040 and volatile organic chemicals (VOCs) listed in 10 CSR 60-4.100(2) and unregulated organic and inorganic chemicals listed in 10 CSR 60-4.110(2).

(B) If the concentration in the composite sample is greater than or equal to 0.0005 mg/l for any organic contaminant listed in 10 CSR 60-4.100(2), then a follow-up sample must be taken within fourteen (14) days at each sampling point included in the composite. These samples must be analyzed for the contaminants which exceeded 0.0005 mg/l in the composite sample. Resampling is not required for unregulated organic and inorganic chemicals listed in 10 CSR 60-4.110(2).

(C) If the concentration in the composite sample detects one (1) or more synthetic organic contaminants listed in subsection (6)(B) of this rule, then a follow-up sample must be taken within fourteen (14) days at each sampling point included in the composite. The follow-up sample must be analyzed for the contaminant(s) detected.

(D) If duplicates of the original sample from each sampling point used in the composite are available, the system may use these duplicates instead of resampling. The duplicate must be analyzed and the results reported to the department within fourteen (14) days of collection.

(E) Compositing Samples Prior to Gas Chromatograph (GC) Analysis.

1. Add five milliliter (5 ml) or equal larger amounts of each sample (up to five (5) samples are allowed) to a twenty-five milliliter (25 ml) glass syringe. Special precautions must be made to maintain zero headspace in the syringe.

2. The samples must be cooled at four degrees Centigrade (4°C) during this step to minimize volatilization losses.

3. Mix well and draw out a five milliliter (5 ml) aliquot for analysis.

4. Follow sample introduction, purging and desorption steps described in the method.

5. If less than five (5) samples are used for compositing, a proportionately small syringe may be used.

(F) Compositing Samples Prior to GC/Mass Spectrophotometer (MS) Analysis.

1. Inject five milliliter (5 ml) or equal larger amounts of each aqueous sample (up to five (5) samples are allowed) into a twenty-five milliliter (25 ml) purging device using the sample introduction technique described in the method.

2. The total volume of the sample in the purging device must be twenty-five milliliters (25 ml).

3. Purge and desorb as described in the method.

(G) For lead and copper monitoring, composite samples from a maximum of five (5) sampling points per composite sample are allowed with prior approval of the department. The reportable value, as required in 10 CSR 60-7.020, for each of the samples in the composite is the concentration detected multiplied by the number of samples composited. If the concentration in the composite sample is less than the detection limit, the reportable value for each of the samples in the composite is the detection limit multiplied by the number of samples composited. If the ninetieth percentile concentration, calculated in accordance with 10 CSR 60-15.010(3)(C), exceeds the lead and copper action level established in 10 CSR 60-15.010(3)(A) and (B), each of the samples from which the composite was derived must be analyzed individually and reported as required in 10 CSR 60-7.020.

(6) Detection Limits.

(A) Detection limits for inorganic chemical analytical methods are the following:

Detection Limits for Inorganic Contaminants

Contaminant	Method	Detection Limit (mg/l)
Antimony	Atomic Absorption—Furnace Technique	0.003
	Atomic Absorption—Platform	0.0008
	Inductively Coupled Plasma—Mass Spectrometry	0.0004
	Atomic Absorption—Gaseous Hydride	0.001
Asbestos	Transmission Electron Microscopy	0.01 million fibers per liter



Contaminant	Method	Detection Limit (mg/l)	Contaminant	Method	Detection Limit (mg/l)	Contaminant	Detection Limit (mg/l)
Barium	Atomic Absorption—Furnace Technique	0.002	Nitrite	Spectrophotometric	0.01	Aroclor 1260	0.0002
	Atomic Absorption—Direct Aspiration	0.1		Automated Cadmium Reduction	0.05	Simazine	0.00007
	Inductively Coupled Plasma	0.002	Manual Cadmium Reduction	0.01	Toxaphene	0.001	
	Inductively Coupled Plasma (EPA method 200.7)	0.001	Ion Chromatography	0.004			
							(C) The detection limit for VOCs is 0.0005 mg/l.
Beryllium	Atomic Absorption—Furnace Technique	0.0002	Selenium	Atomic Absorption—Furnace Technique	0.002		(D) Detection limits for radiological contaminants are the following:
	Atomic Absorption—Platform	0.00002		Atomic Absorption—Gaseous Hydride	0.002		
	Inductively Coupled Plasma	0.0003	Thallium	Atomic Absorption—Furnace Technique	0.001	Radionuclide	Detection Limit
	Inductively Coupled Plasma—Mass Spectrometry	0.0003		Atomic Absorption—Platform	0.0007	Radium 226, 228	1 pCi/l
				Inductively Coupled Plasma—Mass Spectrometry	0.0003	Gross Alpha	3 pCi/l
Cadmium	Atomic Absorption—Furnace Technique	0.0001	(B) Detection limits for SOCs are the following:				
	Inductively Coupled Plasma	0.001	Contaminant	Detection Limit (mg/l)			
Chromium	Atomic Absorption—Furnace Technique	0.001	2,3,7,8-TCDD (Dioxin)	0.000000005			
	Inductively Coupled Plasma	0.007	2,4-D	0.0001			
	Inductively Coupled Plasma (EPA Method 200.7A)	0.001	2,4,5-TP (Silvex)	0.0002			
			Alachlor	0.0002			
			Aldicarb	0.0005			
Copper	All Methods Except Atomic Absorption	0.0002	Aldicarb sulfoxide	0.0005			
	Atomic Absorption With Direct Aspiration	0.0002	Aldicarb sulfone	0.0008			
Cyanide	Distillation, Spectrophotometric (screen)	0.02	Atrazine	0.0001			
	Distillation, Automated Spectrophotometric (screen)	0.005	Benzo(a)pyrene	0.00002			
	Distillation, Selective Electrode (screen)	0.02	Carbofuran	0.0009			
	Distillation, Amenable, Spectrophotometric (free)	0.02	Chlordane	0.0002			
			Dalapon	0.001			
Lead	All Methods	0.001	Dibromochloropropane (DBCP)	0.00002			
Mercury	Manual Cold Vapor Technique	0.0002	Di(2-ethylhexyl)adipate	0.0006			
	Automated Cold Vapor Technique	0.0002	Di(2-ethylhexyl)phthalate	0.0006			
Nickel	Atomic Absorption—Furnace Technique	0.001	Dinoseb	0.0002			
	Atomic Absorption—Platform	0.0006	Diquat	0.0004			
	Inductively Coupled Plasma	0.005	Endothall	0.009			
	Inductively Coupled Plasma—Mass Spectrometry	0.0005	Endrin	0.00001			
			Ethylene dibromide (EDB)	0.00001			
Nitrate	Manual Cadmium Reduction	0.01	Glyphosate	0.006			
	Automated Hydrazine Reduction	0.01	Heptachlor	0.00004			
	Automated Cadmium Reduction	0.05	Heptachlor epoxide	0.00002			
	Ion Selective Electrode	1	Hexachlorobenzene	0.0001			
	Ion Chromatography	0.01	Hexachlorocyclopentadiene	0.0001			
		Lindane	0.00002				
		Methoxychlor	0.0001				
		Oxamyl (Vydate)	0.002				
		Pentachlorophenol	0.00004				
		Picloram	0.0001				
		Polychlorinated biphenyls (PCBs) (as decachlorobiphenyl)	0.0001				
		Aroclor 1016	0.00008				
		Aroclor 1221	0.02				
		Aroclor 1232	0.0005				
		Aroclor 1242	0.0003				
		Aroclor 1248	0.0001				
		Aroclor 1254	0.0001				

(C) The detection limit for VOCs is 0.0005 mg/l.

(D) Detection limits for radiological contaminants are the following:

Radionuclide	Detection Limit
Radium 226, 228	1 pCi/l
Gross Alpha	3 pCi/l
Tritium	1000 pCi/l
Strontium-89	10 pCi/l
Strontium-90	2 pCi/l
Iodine-131	1 pCi/l
Cesium-134	10 pCi/l
Gross Beta	4 pCi/l
Other radionuclides	1/10 of the acceptable limit

AUTHORITY: sections 640.100 and 640.125.1, RSMo Supp. 1998. Original rule filed May 4, 1979, effective Sept. 14, 1979. Amended: Filed April 14, 1981, effective Oct. 11, 1981. Amended: Filed June 2, 1988, effective Aug. 31, 1988. Rescinded and readopted: Filed Dec. 4, 1990, effective July 8, 1991. Rescinded and readopted: Filed March 31, 1992, effective Dec. 3, 1992. Amended: Filed Aug. 4, 1992, effective May 6, 1993. Amended: Filed May 4, 1993, effective Jan. 13, 1994. Amended: Filed Feb. 1, 1996, effective Oct. 30, 1996. Amended: Filed July 1, 1999, effective March 30, 2000.*

**Original authority: 640.100, RSMo 1939, amended 1978, 1981, 1982, 1988, 1989, 1992, 1993, 1995, 1996, 1998, 1999; and 640.125, RSMo 1978, amended 1988.*

10 CSR 60-5.020 Laboratory Certification

PURPOSE: This rule establishes that required analyses must be done by laboratories certified by the department.

(1) For the purpose of determining compliance with this chapter, analytical results will be acceptable only if the samples have been analyzed by a laboratory certified by the department.

(2) To receive approval to conduct analyses for antimony, asbestos, barium, beryllium, cadmium, chromium, cyanide, fluoride, mercury, nickel, nitrate, nitrite, selenium, and thallium the laboratory must—

(A) Analyze performance evaluation samples which include those substances provided by the Environmental Protection Agency (EPA) Environmental Monitoring and Support Laboratory or equivalent samples provided by the department; and

(B) Achieve quantitative results on the analyses that are within the following acceptance limits:

Contaminant	Acceptance Limit
Antimony	±30% at ≥0.006 mg/l
Asbestos	2 standard deviations based on study statistics
Barium	±15% at ≥0.15 mg/l
Beryllium	±15% at ≥0.001 mg/l
Cadmium	±20% at ≥0.002 mg/l
Chromium	±15% at ≥0.01 mg/l
Copper	±10% at ≥0.50 mg/l
Fluoride	±10% at ≥ to 10 mg/l
Lead	±30% at ≥0.005 mg/l
Mercury	±30% at ≥0.0005 mg/l
Nickel	±15% at ≥0.01 mg/l
Nitrate	±10% at ≥0.4 mg/l
Nitrite	±15% at ≥0.4 mg/l
Selenium	±20% at ≥0.01 mg/l
Thallium	±30% at ≥0.002 mg/l

(3) To receive certification to conduct analyses for the contaminants in 10 CSR 60-4.100(2)(A)2.-7. and (B)1.-13., the laboratory must—

(A) Analyze performance evaluation samples which include these substances provided by EPA Environmental Monitoring and Support Laboratory or equivalent samples provided by the department;

(B) Achieve the quantitative acceptance limits in subsections (3)(C) and (D) of this rule for at least eighty percent (80%) of the regulated organic chemicals listed in 10 CSR 60-4.100(2)(A)2.-7. and (B)1.-13.;

(C) Achieve the quantitative results on the analyses performed under subsection (3)(A) of this rule that are within plus or minus twenty percent (±20%) of the actual amount of the substances in the performance evaluation sample when the actual amount is greater than or equal to 0.010 mg/l;

(D) Achieve quantitative results on the analyses performed under subsection (3)(A) of this rule that are within plus or minus forty percent (±40%) of the actual amount of the substances in the performance evaluation sample when the actual amount is less than 0.010 mg/l; and

(E) Achieve a method detection limit of 0.0005 mg/l.

(4) To receive certification for vinyl chloride, the laboratory must—

(A) Analyze performance evaluation samples provided by EPA Environmental Monitoring and Support Laboratory or equivalent samples provided by the department;

(B) Achieve quantitative results on the analyses performed under subsection (4)(A) of this rule that are within plus or minus forty percent (±40%) of the actual amount of vinyl chloride in the performance evaluation sample;

(C) Achieve a method detection limit of 0.0005 mg/l; and

(D) Obtain certification for the contaminants listed in 10 CSR 60-4.100(2)(A)2.-7. and (B)1.-13.

(5) To receive certification to conduct analyses for the contaminants in 10 CSR 60-4.040(1), the laboratory must—

(A) Analyze performance evaluation samples which include those substances provided by EPA Environmental Monitoring and Support Laboratory or equivalent samples provided by the department.

(B) Achieve quantitative results on the analyses that are within the following acceptance limits:

Contaminant	Acceptance Limit
	(percent)
2,3,7,8-TCDD (Dioxin)	2 standard
2,4-D	±50
2,4,5-TP	±50
Alachlor	±45
Aldicarb	2 standard deviations
Aldicarb sulfoxide	2 standard deviations
Aldicarb sulfone	2 standard deviations
Atrazine	±45
Benzo(a)pyrene	2 standard deviations
Carbofuran	±45
Chlordane	±45
Dalapon	2 standard deviations
Dibromochloropropane	±40
Di(2-ethylhexyl)adipate	2 standard deviations
Di(2-ethylhexyl)phthalate	2 standard deviations
Dinoseb	2 standard deviations
Diquat	2 standard deviations
Endothall	2 standard deviations
Endrin	±45

Ethylene dibromide	±40
Glyphosate	2 standard deviations
Heptachlor	±45
Heptachlor epoxide	±45
Hexachlorobenzene	2 standard deviations
Hexachlorocyclopentadiene	2 standard deviations
Lindane	±45
Methoxychlor	±45
Oxamyl	2 standard deviations
Polychlorinated biphenyls (PCBs) (as decachlorobiphenyl)	0—200
Picloram	2 standard deviations
Simazine	2 standard deviations
Toxaphene	±45
Pentachlorophenol	±50

(6) To receive approval to conduct analyses for copper and lead, the laboratory must—

(A) Analyze performance evaluation samples which include those substances provided by EPA Environmental Monitoring and Support Laboratory or equivalent samples provided by the department;

(B) Achieve quantitative acceptance limits for copper plus or minus ten percent (±10%) of the actual amount in the performance evaluation sample when the actual amount is greater than or equal to 0.050 mg/l; lead plus or minus thirty percent (±30%) of the actual amount in the performance evaluation sample when the actual amount is greater than or equal to 0.005 mg/l; and

(C) Achieve a method detection limit of 0.001 mg/l.

(7) Analysis for disinfection byproducts must be conducted by laboratories that have received certification by the department except that a party approved by the department must measure daily chlorite samples at the entrance to the distribution system. To receive certification to conduct analyses for the TTHM, HAA5, bromate and chlorite, the laboratory must carry out annual analyses of performance evaluation (PE) samples approved by the department. In these analyses of PE samples, the laboratory must achieve quantitative results within the acceptance limit on a minimum of eighty percent (80%) of the analytes included in each PE sample. The acceptance limit is defined as the ninety-five percent (95%) confidence interval calculated around the mean of the PE study data between a maximum and minimum acceptance limit of plus or minus fifty percent

($\pm 50\%$) and plus or minus fifteen percent ($\pm 15\%$) of the study mean.

(8) The department has the authority to allow the use of previously collected monitoring data for purposes of monitoring, if the data were collected and analyzed in accordance with the requirements of this rule.

(9) All lead levels measured between the Practical Quantification Level (PQL) and Method Detection Limit (MDL) must be either reported as measured or they can be reported as one-half ($1/2$) the PQL (0.0025 mg/l). All levels below the lead MDL must be reported as zero (0).

(10) All copper levels measured between the PQL and the MDL must be either reported as measured or they can be reported as one-half ($1/2$) the PQL (0.015 mg/l). All levels below the copper MDL must be reported as zero (0).

(11) Operational monitoring measurements required by 10 CSR 60-4.080(3) shall be performed on-site by persons acceptable to the department.

(12) The department will consider acceptance of analytical results from out-of-state laboratories upon written request.

AUTHORITY: section 640.100, RSMo Supp. 1999. Original rule filed May 4, 1979, effective Sept. 14, 1979. Rescinded and readopted: Filed March 31, 1992, effective Dec. 3, 1992. Amended: Filed Aug. 4, 1992, effective May 6, 1993. Amended: Filed May 4, 1993, effective Jan. 13, 1994. Amended: Filed Feb. 1, 1996, effective Oct. 30, 1996. Amended: Filed Dec. 15, 1999, effective Sept. 1, 2000.*

**Original authority: 640.100, RSMo 1939, amended 1978, 1981, 1982, 1988, 1989, 1992, 1993, 1995, 1996, 1998, 1999.*